



CERTIFICATE No 01-24 (revised certificate No. 09-17)
SET OF REFERENCE MATERIALS OF SOLID FUEL AND ASH
for thermodynamic, chemical and technological properties

Status

The SF and SFA Reference Materials comply with the ISO Guide 35 definition of the Reference Material.

Intended for quality control and validation of methods for gross calorific content measurement, elemental analysis of C, N, H, S content and determination of the conventional of volatile matter and ash content. For this, the values in the set are evenly distributed over the entire application ranges.

They may not substitute the CRM in establishing traceability of values.

The individual RMs of natural and processed fuels ranging from brown coal over coke to anthracite make it possible to verify the matrix-match in testing the particular fuel type. One ash standard SFA-01-14 was added to this set.

User instruction: all stated values are assigned to a dry basis, thus the moisture content (measured after one hour drying at 105°C) should be determined concurrently (within 24 hours) and the other measurement results corrected accordingly. The proximate moisture content of the RM is 0,2-4% wt. The measurements are valid provided the respective standard methods (annex) are applied and the **expiration time (3 years from the first opening of vial)** and a minimum sample intake are observed. This is 0,5 g for calorific content, 0,07 g for the elemental analysis and 1 g for the conventional values. The vial lid should be replaced immediately after each sample intake.

Storage in a dry environment with an ambient temperature below 25°C is recommended. There are no safety hazards in the proper storage and use of CRM.

Candidate materials, selected to match the target properties, were milled and sieved to separate the fraction with over 90 wt.% between 0,05 and 0,2 mm grain size and matured for at least two years, stored at the regular conditions. They were ultimately homogenized again and adjusted to the working vials.

Supplied in a set or as the individual expedition units consisting 50g of RM each.

The vial label allows marking the first opening date to facilitate the expiration time control.

Manufactured and characterised in a strict compliance with ISO Guides 34, 35 by:
producer SPL Bohumín.

Produced by: SPL, the authorised producer of CRM for the Czech Metrology Institute and the provider of the interlaboratory Proficiency Testing accredited by the Czech Accreditation Institute for the chemical analysis of metallic materials, in a strict compliance with ISO/IEC 17025, 17043 and in particular with ISO Guide 34.

Certified values on the reverse side are based on an international collaborative study. They are means of at least eight laboratory sets of results, accepted after their technical assessment.

One laboratory obtained two randomly selected vials and reported five parallel independent results for each. The results were processed by two-way-nested ANOVA (ISO Guide 35 (2006), annex A2), which provides, besides of mean value and its uncertainty estimates, a possible within-vials homogeneity detection.

The certified values are rounded to the same digit, as their uncertainty statement.

Laboratories listed below participated in the collaborative study. They were accredited and/or they demonstrated compliance with ISO/IEC 17025 during the respective measurement.

Participating laboratories:

AmpluServis, Ostrava, Czechia

ArcelorMittal, Ostrava, Czechia

Coal Services, Most, Czechia

Deza, Valašské Meziříčí, Czechia

Elektrárna Dětmarovice, Dětmarovice, Czechia

Energopomiar, Gliwice, Poland

Enviform, Třinec, Czechia

Instytut Chemicznej Przeróbki Węgla, Zabrze, Poland

Leco ETC., Praha, Czechia

OKD, Karviná, Czechia

OKK Koksovny, Ostrava, Czechia

Orgrez, Most, Czechia

SES Inspekt, Tlmače, Slovakia

Štátny geologický ústav Dionýza Štúra, Spišská Nová Ves, Slovakia

Vítkovice Testing Center, Ostrava, Czechia

VŠB, Ostrava, Czechia

VVUÚ, Ostrava, Czechia

Issued in January 2024

Responsible person
Martin Bogumský

SPL-LABMAT s.r.o.
1. máje 432
735 31 Bohumín, CZ
IČO: 06480870, DIČ: CZ06480870
www.spl-labmat.cz
e-mail: info@spl-labmat.cz